



B-BK-60212-0446/20

ORIGINAL

HYGIENIC CERTIFICATE

NATIONAL INSTITUTE OF PUBLIC HEALTH - NATIONAL INSTITUTE OF HYGIENE

Product: **Lamps UV-C EXTERYA 1, 2, 3, 4, 5, 6, 7, 8, 9, 10**

Containing: powder-coated aluminium and other materials according to manufacturer's documentation

Destined: for use in doctors' surgeries, hotels, kindergartens, schools, nurseries, pharmacies, health centres, beauty parlours, housing and commercial premises

The above-named product is acceptable according to hygienic criteria with the following conditions:

The application and manufacture of the product must comply with the regulations applicable to the facility in which it is used. Installation and operation in accordance with the manufacturer's instructions.

During the operation of direct-acting UV-C lamps, no persons are allowed in the room where the device is installed.

The presence of people in the room is only possible during operation of the UV-C flow-through lamps.


Hygienic certificate does not apply to technical parameters, utility value and allergenic properties of the product

Producer:

EXTERYA Sp. z o. o.
05-250 Słupno
Zakład Produkcyjny, ul. Żeromskiego 39A

This certificate issued for:

EXTERYA Sp. z o. o.
02-634 Warszawa
ul. Raclawicka 114



The certificate may be corrected or cancelled after appropriate motivation. The certificate loses its validity after 2025.11.27 or in the case of changes in composition or in technology of production.

The date of issue of the certificate: 27th November 2020

Head of the Department of Environmental
Health Safety

Jolanta Solecka
dr hab. Jolanta Solecka, prof. NIZP-PZH